



Test Report

Company Name & Address Honsin Apparel Sdh. Bhd.

531, Batu 2 ½, Jalan Kluang,

83000, Batu Pahat,

Johor Darul Takzim, Malaysia.

Product/Model(s) ProXmask Filter

Standard/Specification(s) Breathability Test (Differential Pressure)

EN 14683:2019

Date of Test 24 September 2020

Test Report Number 7191244646-EEC20-02-WBH

This report is made on the basis of information and data from original report results.

Prepared by

Dr Zetty Azalea R&D Chemist



1. Breathability Test (Differential Pressure)

Test description:

TEST REPORT

BS EN 14683:19 Medical face masks – Requirement and test methods
Clause 5.2.3 Breathability/ Annex C Method for determination of breathability (differential pressure)

Airflow: 8 L/min Tested area: 4.9 cm²

Number of tested specimen: 5 Number of test area of specimen: 5

Sample description:

ProXmask filter

Result:

Test specimen	Average differential pressure (Δ P)	Requirement
	for tested area (Pa/cm²)	
1	56	BS EN 14683:2019
2	57	
3	58	Type I <40 Pa/cm²
4	56	Type II <40 Pa/cm ²
5	57	Type IIR <60 Pa/cm²